

ARCHITECT INSTRUMENT SETTING

Angiotensin Converting Enzyme - ACE (BSBE)

Method : Colorimetric

<input checked="" type="radio"/> Genral	<input type="radio"/> Calibration	<input type="radio"/> Smartwash	<input type="radio"/> Result	<input type="radio"/> Interpretation	
Assay :	<input type="text" value="ACE"/>	Type :	<input type="text" value="Photometric"/>	Version :	<input type="text" value="1"/>
Number :	<input type="text"/>	Assay :	<input type="text" value="Enabled"/>	Date :	<input type="text"/>
				Time :	<input type="text"/>
				Operator :	<input type="text" value="ADMIN"/>

<input checked="" type="radio"/> Reaction Defination	<input type="radio"/> Reagent / Sample	<input type="radio"/> Vallidity Check				
Reaction mode :	<input type="text" value="Rate down"/>					
		Read Time				
Wavelength	<table style="width: 100%; border: none;"> <tr> <td style="text-align: center;">Primary</td> <td style="text-align: center;">Secondary</td> </tr> <tr> <td style="text-align: center;"><input type="text" value="340 / 404"/></td> <td></td> </tr> </table>	Primary	Secondary	<input type="text" value="340 / 404"/>		Main : <input type="text" value="5 - 28"/>
Primary	Secondary					
<input type="text" value="340 / 404"/>						
Last Required read	<input type="text" value="28"/>	Flex : <input type="text" value="0-0"/>				
Absorbance range	<input type="text" value="0.0000 - 3.0000"/>	Color correction : <input type="text" value="0-0"/>				
Sample blank type	<input type="text" value="None"/>	Blank : <input type="text"/>				

<input type="radio"/> Reaction Defination	<input checked="" type="radio"/> Reagent / Sample	<input type="radio"/> Vallidity Check					
Reagent :	<input type="text" value="ACE"/>	Reagent Volume :	<table style="width: 100%; border: none;"> <tr> <td style="text-align: center;">R1</td> <td style="text-align: center;">R2</td> </tr> <tr> <td style="text-align: center;"><input type="text" value="200"/></td> <td><input type="text"/></td> </tr> </table>	R1	R2	<input type="text" value="200"/>	<input type="text"/>
R1	R2						
<input type="text" value="200"/>	<input type="text"/>						
Diluent	<input type="text" value="Saline"/>	Water volume :	<input type="text" value="0"/>				
Diluent dispense mode :	<input type="text" value="Type0"/>	Dispense mode :	<table style="width: 100%; border: none;"> <tr> <td style="text-align: center;">Type0</td> <td><input type="text"/></td> </tr> </table>	Type0	<input type="text"/>		
Type0	<input type="text"/>						
Dilution name	Sample	Diluted Sample	Diluent	Water	Dilution Factor	Default dilution	
STANDARD	<input type="text" value="17"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="1:0"/>	<input checked="" type="radio"/>	
1:2	<input type="text" value="35"/>	<input type="text" value="15"/>	<input type="text" value="65"/>	<input type="text" value="0"/>	<input type="text" value="1:2"/>	<input type="radio"/>	

<input type="radio"/> Reaction Defination	<input type="radio"/> Reagent / Sample	<input checked="" type="radio"/> Vallidity Check	
Reaction check type :	<input type="text" value="None"/>	Read time B range :	<input type="text"/>
Read time A range :	<input type="text"/>	Minimum absorbance :	<input type="text"/>
Calculation limit :	<input type="text"/>	Rate linearity % :	<input type="text" value="0"/>

<input type="radio"/> Genral	<input checked="" type="radio"/> Calibration	<input type="radio"/> Smartwash	<input type="radio"/> Result	<input type="radio"/> Interpretation	
Assay :	<input type="text" value="ACE"/>	Assay Numbe :	<input type="text"/>	Date :	<input type="text"/>
Calibration method :	<input type="text" value="Linear"/>	Time :	<input type="text"/>	Operator :	<input type="text" value="ADMIN"/>

<input checked="" type="radio"/> Calibrators	<input type="radio"/> Volumes	<input type="radio"/> Intervals	<input type="radio"/> Validity checks
Calibrator Set :	Calibrator level :	Concentration :	
<input type="text" value="ACE"/>	Blank : <input type="text" value="Water"/>	<input type="text" value="0.00"/>	
	Cal 1 : <input type="text" value="ACE CAL 1"/>		
Replicates :	<input type="text" value="2"/>		

<input type="radio"/> Calibrators	<input checked="" type="radio"/> Volumes	<input type="radio"/> Intervals	<input type="radio"/> Validity checks		
Calibrator :	<input type="text" value="ACE Cal"/>				
Blank	Calibrator Level	Sample	Diluted sample	Diluent	Water
	Water	17	0.0	0	0
Cal 1	ACE Cal	17	0.0	0	0

<input type="radio"/> Genral	<input type="radio"/> Calibration	<input checked="" type="radio"/> Smartwash	<input type="radio"/> Result	<input type="radio"/> Interpretation	
Componennt	Reagent / Assay	Wash	Volume	Replicates	Wash Protocol
Sample Prob		Water			Madium Wash
Reagent Probe (R1)	All	Water	345	2	

<input type="radio"/> Genral	<input type="radio"/> Calibration	<input type="radio"/> Smartwash	<input checked="" type="radio"/> Result	<input type="radio"/> Interpretation
Linearty range :		0.00 - 150.00		
Flag range specificaation :				
Gender	Age	Normal Range	External range	
Male	10-90 Years	12.00 - 68.00		

Result units			
Result Concentration unit :	<input type="text" value="U/L"/>	Correlation factor :	<input type="text" value="1.0000"/>
Result decimal places :	<input type="text" value="1"/>	Intercept :	<input type="text" value="0.0000"/>