

ARCHITECT INSTRUMENT SETTING

OXALATE (BEN)

Method : Quantitative UV Assay

 Genral Calibration Smartwash Result InterpretationAssay : Type : Version : Number : Assay : Date : Time : Operator : Reaction Defination Reagent / Sample Vallidity CheckReaction mode :

Read Time

Wavelength Main : Last Required read Flex : Absorbance range Color correction : Sample blank type Blank : Reaction Defination Reagent / Sample Vallidity CheckReagent :

Reagent Volume :

R1	R2
<input type="text" value="180"/>	<input type="text" value="36"/>

Diluent

Water volume :

Diluent dispense mode :

Dispense mode :

Type0	Type0
<input type="text" value="Type0"/>	<input type="text" value="Type0"/>

Dilution name	Sample	Diluted Sample	Diluent	Water	Dilution Factor	Default dilution
STANDARD	<input type="text" value="9"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="1:1"/>	<input checked="" type="radio"/>
1:2	<input type="text" value="10"/>	<input type="text" value="9"/>	<input type="text" value="200"/>	<input type="text" value="0"/>	<input type="text" value="1:2.00"/>	<input type="radio"/>

 Reaction Defination Reagent / Sample Vallidity CheckReaction check type : Read time B range : Read time A range : Minimum absorbance : Calculation limit :

Rate linearity % :

<input type="radio"/> Genral	<input checked="" type="radio"/> Calibration	<input type="radio"/> Smartwash	<input type="radio"/> Result	<input type="radio"/> Interpretation	
Assay :	<input type="text" value="Oxalate"/>	Assay Numbe :	<input type="text"/>	Date :	<input type="text"/>
Calibration method :	<input type="text" value="Linear"/>	Time :	<input type="text"/>	Operator :	<input type="text" value="ADMIN"/>

<input checked="" type="radio"/> Calibrators	<input type="radio"/> Volumes	<input type="radio"/> Intervals	<input type="radio"/> Validity checks		
Calibrator Set :	<input type="text" value="Oxalate CAL"/>	Calibrator level :	<input type="text" value="Water"/>	Concentration :	<input type="text" value="0.00"/>
		Blank :	<input type="text" value="Oxalate Cal"/>		
		Cal 1 :	<input type="text" value="Oxalate Cal"/>		
		Replicates :	<input type="text" value="1"/>		

<input type="radio"/> Calibrators	<input checked="" type="radio"/> Volumes	<input type="radio"/> Intervals	<input type="radio"/> Validity checks		
Calibrator :	<input type="text" value="Oxalate Cal"/>				
Blank	Calibrator Level	Sample	Diluted sample	Diluent	Water
	Water	9	0.0	0	0
Cal 1	Oxalate Cal	9	0.0	0	0

<input type="radio"/> Genral	<input type="radio"/> Calibration	<input checked="" type="radio"/> Smartwash	<input type="radio"/> Result	<input type="radio"/> Interpretation	
Componennt	Reagent / Assay	Wash	Volume	Replicates	Wash Protocol

<input type="radio"/> Genral	<input type="radio"/> Calibration	<input type="radio"/> Smartwash	<input checked="" type="radio"/> Result	<input type="radio"/> Interpretation	
Linearty range :		3.00 - 400.00			
Flag range specificaation :					
Gender	Age	Normal Range	External range		
Either	0-100 Years	0.20 - 6.10			

Result units			
Result Concentration unit :	<input type="text" value="mg/dL"/>	Correlation factor :	<input type="text" value="1.0000"/>
Result decimal places :	<input type="text" value="2"/>	Intercept :	<input type="text" value="0.0000"/>