

DADE DIMENSION PARAMETER SETTINGS

Gcell

CA

Cat. No: GB400E-GS401E

Channel: *	Name:	CA	Mode:	ABSORBENCE	Std. curve:	LINEAR
Delivery						
	Time	Component 1	Component 2	Component 3	Chase	Mix
R1	-60.0 sec	(A)200 µl	0 µl	0 µl	20 µl	NONE
S1	0.0 sec	3 µl	0 µl	0 µl	20 µl	MODERATE
R2	sec	(B) µl	0 µl	0 µl	0 µl	
R3	0 sec	0 µl	0 µl	0 µl	NONE µl	
Photometry						
	Time	Cartridge configuration				
P1	-30 sec	Component	(A)	(A)	(A)	(A)
P2	300 sec	Number of tests	16	16	16	16
P3	sec	Well life hours	720	720	720	720
P4	sec	On board life 720 hours	Calibration 2160 hours			
Calculation						
Mode	END POINT	Measuring filter	600 nm	Blanking filter		nm
P1	Time -30 sec	Dilution	-1.0	IOD	0	(Inactive)
P2	Time 300 sec	Dilution	1.0	FOD	0	(Inactive)
	1	2	3	4	5	6
	*	*	*			
Decimal	2					
Units	mg/dL					
Range	8.10-10.4					
Assay Range	21					